



P O BOX 15666, LYTTTELTON 0140 scales@richterscale.co.za www.richterscale.co.za
 TEL: +27(12) 386 2060/1 FAX: +27(12) 386 0345

QD : 12 / 01 / 2007

VERIFICATION CERTIFICATE

Certificate No. RSC 3612

Customers Name: <u>Makeean Polymers</u>	Test Address: <u>236 Propshaft Road Samcor Park, Waltloo</u>	Usage Address: <u>236 Propshaft Road Samcor Park, Pretoria</u>
--------------------------------------------	---------------------------------------------------------------------	-----------------------------------------------------------------------

SCALE DETAILS

Make: Jadever Model: JIK6 CAB Capacity: 1000 kg Division Size : 0.5 kg
 Serial No. IK1120111N0960 S.A. Approval No. 1402 I/O Location: Richter Scale Workshop
 Observations made during the Preliminary Examination: Good
 Test Conditions e.g. environment: Good
 Status Prior to Verification: New: Repaired: In-Trade-Use: Rejected:

PERMISSABLE TOLERANCES

0	Kg To	250	Kg	0,25	Kg
250	Kg To	1000	Kg	0,5	Kg
	Kg To		Kg		Kg
	Kg To		Kg		Kg
	Kg To		Kg		Kg
DISCRIMINATION / SENSITIVITY TOL:					0,6 kg

ACCURACY TEST

+Run	Result	-Run	Result
0	kg	✓	
10	kg	✓	
250	kg	✓	
500	kg	✓	
750	kg	✓	
1000	kg	✓	

Load Applied

Ref S/P _____ kg @ _____ kg
 Make-up mass _____ kg
 Make-up mass _____ kg
 Ref S/P 5,5/6 kg @ 0,25 kg

ECCENTRIC TEST (Not for Counter and SSI Scales)

No of load points: 4
505,5/6 kg @ 0,25 kg
 ✓ ✓ _____ _____
 ✓ ✓ _____ _____

Variation Test (END MIDDLE END)

> _____ > _____ > _____
 < _____ < _____ < _____
 > _____ > _____ > _____

Price Computation _____ Tare
 Repeatability Printer _____
 Blanking Segment Check
 Lateral Test _____ Spirit Level _____
 Discrimination: Sensitivity _____
 Back Indication : _____ Auto-Zero Tracking _____

✓ = CORRECT X = INCORRECT

A line to be put through the certificate sections which are not applicable. Page 1/1

VERIFICATION CERTIFICATE

This instrument was tested and found to comply in all respects of the Trade Metrology Act, 1973 (Act 77 of 1973) and may be used for a prescribed purpose as intended by the act.

Expiry date of verification certificate: 25/06/2014

STANDARDS

The masspieces used for this verification are traceable to national standards through an approved SANAS laboratory.

Calibration Cert No. ECS/12 /530 - ECS/11/450

Date of Calibration: 17-20/01/2012 - 9/9/2011

Expiry Date of Calibration: 16/01/2013 - 8/9/2012

I.D. of Equipment: BoxR0678/Black500kg

Own or Loan Equipment: Own

VERIFICATION OFFICER

Name and Certificate No. P. Wolfswinkel M747

Date: 25/06/2012

Seal / Stamp No. Richter Scale Co 3

Verification Procedure RS/VP/01

Signature: _____



SANAS REG NO : LTF 0103